

Accreditation Scope

**Aries Oil Field Services LLC, NAL 122
Calibration Laboratory, (ISO/IEC 17025:2017)**

Mussafah, Abu Dhabi, UAE

Issue Date: 19-08-2020

Expiry Date: 12-12-2021

Issue No: 02

Calibration Field/ Quantity/ Property	Measurand / Equipment	Measuring Range	CMC (k=2)	Calibration Method (Standard/ Internal Procedure)	Permanent lab (P) / Client-site (S)
Pressure	Gauge Pressure (Hydraulic)	(5 to 60) bar	0.1 bar	DKD-R-6-1:2014 QHSE/P11/CAL/JP01/PG:2019	P
		(> 60 to 700) bar	0.2 bar	DKD-R-6-1:2014 QHSE/P11/CAL/JP01/PG:2019	P
	Negative Gauge Pressure	(-0.8 to -0.1) bar	0.06 bar	DKD-R-6-1:2014 QHSE/P11/CAL/JP02/VG:2019	P
Dimensional	Vernier Caliper – External dimensions only (Digital / Analog)	(0 to 300) mm	30 µm	QHSE/P11/CAL/JP08/VC:2019	P
		(300 to 600) mm	40 µm	QHSE/P11/CAL/JP08/VC:2019	P
END					